



### **Complete Selection System**

Includes templates and models to aid in implant selection

UPOR

Ordering Instructions

1. Using both the clear template and color template, pick out the closest matching ear.

2. When placing the order, be sure to order the catalog number for the correct side.

10. 4651 m ancies she fally formed ear, and the correct side.

10. 5600, will need to be referred.

11 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

12 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

13 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

14 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

15 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

16 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

17 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

18 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.


19 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our Platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our platent Specific solutions.

19 you can't find an ear to match your patients's needs, contact us to discuss our platent Specific solutions.

Match Size and
Contour Details
to Your Patient's
Anatomy and
Reconstructive
Requirements



SU-POR

Implants for Life  $^{\scriptscriptstyle TM}$ 

### Available in 24 Unique Designs

### PLACING YOUR ORDER

When placing the order, be sure to order the catalog number for the side requiring reconstruction. Utilize the convenient Ear Legend located on the underside of the sizing kit.

Submit all order requests by:

email: sales@poriferous.com phone: (877) 631 - 1954 (Toll-Free)



Instructions For Use R & Helpful Videos

www.poriferous.com

If you can't find an ear to match your patient's needs, contact us to discuss our Patient-Specific solutions

## Groundbreaking Product That Offers A New Level of Efficiency for Reconstructive Surgery


# 



The implant's open-pore structure enables integration with the body.

TRUSTED TECHNOLOGY WITH OVER 40 YEARS OF CLINICAL POST-OPERATIVE DATA.

With its innovative technology, an integrated Su-Por implant has its own blood supply, thanks to vascular fascia, making it an ideal option for full-thickness skin grafts.



The Su-Por implant promotes rapid tissue growth, becoming fully mucosalized in three to four weeks. Within three weeks of implantation, the implant typically integrates with fibrovascular tissue, and bone integration occurs after six weeks in recently cut areas.

# IMPLANT PREPARATION

### SOAKING THE IMPLANTS WITH ANTIBIOTIC SOLUTION

- 1. Prophylactic measure against any contaminates introduced during surgery
- 2. Wetting implant allows for better, more complete tissue integration
- 3. Improves fibrovascular and bone ingrowth into implants pores



### **CUTTING & CONTOURING**

Su-Por Surgical Implants are readily cut with surgical instruments. The implants may be carved in the sterile field or directly in vivo, to meet the needs of your patients. Scalpel, scissors, or burr may be used to achieve the desired contour and detail. After adjustments are made, rinse the implant in sterile saline to remove any particulate or debris.

### **ANTIBIOTIC SELECTION**

The solution is dependent on patient tolerance and antibiotic need. There are no known drug incompatibilities with porous polyethylene.







# Left Right **4600/4601**

**57mm** (Height)

32mm (Width)

18mm (Base + Projection)





# Left Right **4602/4603**

**57mm** (Height)

36mm (Width)


26mm (Base + Projection)



# Left Right 4604/4605

58mm (Height)

38mm (Width)






Left Right 4606/4607

**56mm** (Height)

38mm (Width)

27mm (Base + Projection)








Left Right **4610/4611** 

56mm (Height)

35mm (Width)







# Left Right 4612/4613

55mm (Height)

35mm (Width)

31mm (Base + Projection)





# Left Right **4614/4615**

**58mm** (Height)

36mm (Width)

32mm (Base + Projection)



# Left Right **4616/4617**

53mm (Height)

35mm (Width)






# Left Right 4619

**56mm** (Height)

33mm (Width)











Left Right 4624/4625

55mm (Height)

40mm (Width)

24mm (Base + Projection)





Left Right **4626/4627** 

**56mm** (Height)

37mm (Width)

23mm (Base + Projection)



Left Right **4628/4629** 

60mm (Height)

40mm (Width)









Left Right 4630/4631

**59mm** (Height)

37mm (Width)

31mm (Base + Projection)





Left Right **4632/4633** 

**53mm** (Height)

36mm (Width)


**28mm** (Base + Projection)



Left Right **4634/4635** 

50mm (Height)

33mm (Width)







Left Right 4636/4637

**60mm** (Height)

38mm (Width)

30mm (Base + Projection)





Left Right **4638/4639** 

**63mm** (Height)

38mm (Width)


31mm (Base + Projection)



Left Right **4640/4641** 

52mm (Height)

35mm (Width)







Left Right **4642/4643** 

**57mm** (Height)

38mm (Width)

23mm (Base + Projection)





Left Right **4644/4645** 

**55mm** (Height)

34mm (Width)


**21mm** (Base + Projection)



Left Right **4646/4647** 

56mm (Height)

34mm (Width)



All products shippd by Poriferous, LLC are provided STERILE. DO NOT USE any implant if the package is open, damaged or wet. All STERILE items are considered single use and can not be sterilized.

Surgeons should consult the package insert for use and precautions. For additional information, please contact your local Su-Por Surgical Implant distributor.

All products manufactured and distributed by Poriferous, LLC are latex free.



### **CONTACT US**

www.poriferous.com T: (877) 631 - 1954 (Toll Free) F: (770) 683 - 7459



Copyright 2015. All rights reserved. The Su-Por service mark is owned by or used under the license and authority if Poriferous, LLC. Su-Por is a registered trademark of Poriferous, LLC. 220124 REV C

